

Lighting Passport™

CALIBRATION CERTIFICATE

Model	ALP-01
Serial No.	0005130010830
Calibration Date	2013.08.30
Ambient Temperature	(23±1)°C
Relative Humidity	(45±10)%

Item	Standard Value	Check	Δ	
Illuminance (lux)	405	405	0%	
Chromaticity Coordinates	x	0.4342	0.4350	0.0008
	y	0.4021	0.4022	0.0001
Specification	Δx,y<±0.002, Δlux<±3%			

Calibration Descriptions

1. Location of Test

- This test was performed at 9F., No.43, Fuxing Rd., Xindian Dist., New Taipei City 23150, Taiwan (R.O.C.) , Asensetek Lab.

2. Calibration System includes

- Low pressure line lamp for wavelength calibration
- NIST Traceable Lighting Source System and Spectrophotometer for photometric calibration

Calibration Interval

If the system is handled with care, recalibration is recommended to be made once a year, at the discretion of the user.

QC: 李新峰

Supervisor: 王培全



Address : 8F., No.43, Fuxing Rd., Xindian Dist., New Taipei City 23150, Taiwan (R.O.C.)

Phone : +886-2-2218-5577

FAX : +886-2-2218-0656

E-mail : service@asensetek.com

This certificate shall not be reproduced except in full, without the written approval of Asensetek Inc.